

5.5.6 Calibration overviews

5.5.6.1 Cold/Hot calibrations

Adj.No	Adj	Cold	Hot
10	1 MHz f.s.	x	-
11	1 uA DC zero (open)	x	x
12	100 uA DC zero (open)	x	x
13	20 nF zero (open)	x	-
14	200 Ω FW zero	x	x
15	2 k Ω FW 1 k Ω	x	x
16	20 k Ω FW 10 k Ω	x	x
17	TDC 0 Ω	x	x
18	TDC 100 Ω	x	x
19	200 M Ω TW 30 M Ω	x	x
20	1 uADC 1 uA	x	x
21	10 uADC 10 uA	x	x
22	1 ADC 1 A	x	x
23	200 m VDC zero	x	x
24	2 VDC zero	x	x
25	20 VDC zero	x	x
26	200 VDC zero	x	x
27	2 VPU zero	x	x
28	2 VPL zero	x	x
29	200 mVDC 200 mV	x	x
30	2 VDC 2 V	x	x
31	20 VDC 20 V	x	x
32	20 VDC -20 V	x	-
33	2 VPU 2 V	x	x
34	2 VPL -2 V	x	x
35	200 mVAC 200 mV	x	-
36	2-VAC 2 V	x	x
37	20 VAC 20 V	x	-
38	200 k Ω 100 k Ω	x	x
39	2 M Ω 1 M Ω	x	x
40	20 M Ω 10 M Ω	x	x
41	1 mADC 1 mA	x	x
42	100 mADC 100 mA	x	x
43	1 mAAC 1 mA	x	x
44	200 VAC 200 V	x	-
45	2000 VAC 600 V	x	-
46	200 VDC 200 V	x	-
47	2000 VDC 1000 V	x	-